

# Eddyfi Magnifi® 4.8R19 Release Notes

Released on: December 16<sup>th</sup>, 2020

Note: Magnifi 4.8R19 is identical to Magnifi 4.8R18. Revision number has incremented due to a minor modification in Magnifi GO 4.8R19 only.

## Minimum System Requirements

- **Operating systems:** Windows® 8.1 and Windows 10 version 1607 (Anniversary Update)
  - **Note:** The software is tested and optimized for most major language packs available on the Windows suite.
- **Processor:** Core i5 (or equivalent)
- **Memory:** 8 GB (recommended: 16GB recommended for very large tube maps)
- **Disk space:** 500 GB
- **Recommended network:** Built-in network card (USB-to-network adapter also acceptable)
- **Display:** 13" (recommended: 15")
  - **Display scale** should be set at 100% in Windows Preferences
- **Resolution:** 1366 × 768 pixels (recommended resolution: 1920 × 1080 pixels)
  - For extensive analysis purposes, we recommend using an additional external monitor, 22" or larger with a minimum resolution of 1920 × 1080 pixels
- **Administrator rights :** User must have Administrator permissions on the computer to install and use Magnifi

## Firmware

Included in this release of **Magnifi** is the following Package:

### Eddyfi Ectane® 2

- Version: 2.1R6  
Update your firmware the first time you connect to Ectane 2.

### Ectane

- Version: 1.8R5.1  
This is the same version as Magnifi 3.5R14

## New Features and Improvements

### Generic

- New "Get Assistance" feature for improved troubleshooting/support experience

## Tubing Applications

- New RFT dual receiver probe added to wizard

## Modifications to Existing Features

- After creation of a new tubing setup with Wizard, strip charts are automatically configured to display the vertical component of the signal. Also, they automatically display the channel names and frequencies.

## Dropped Features

- “Save data” button in the Sharck ribbon of Magnifi R removed.

## Resolved Issues

- Iris:
  - Mini cursors synchronize between the projection views
  - Displayed values in infocenter and reported values synchronized
  - WT infocenter displays the right value when the D-scan is hidden
  - Size and WT values shown in cases where backwall echo is missed
  - WT and WL value in projection views refresh when moving cursor in C-scan view
- Supports large window size for median filters in Advanced Process Channels
- Closing data processing window executed properly
- No data file duplication when adding indication or renaming file
- Some actions (add a defect, tag as NDD, screen capture, etc.) are now fixed to act on the first click of the button
- Lift-off assistant feature no longer displays intermittent error messages
- Best fit function works without error notifications
- Software update through the “Check for update” menu functional
- When adding voltage plane to layout, associated control buttons are automatically displayed
- Calibration values correctly applied during C-scan calibration in EDF measurement convention
- Access to Sharck™ wizard locked when Sharck probe connected to avoid faulty setups being generated
- Indication boxes in C-scan disappear between two separate acquisitions
- Maximum scan size not reduced in raster scan mode with Spyne

## Known Issues, Limitations, and Restrictions

- Data saving not available when in Cal.check mode and Lift-off assistant mode
- 3D C-scan display issues (related to Windows 10 drivers)
- 2D and 3D polar C-scans display issues (related to Windows 10 drivers)

- Loss of signal amplitude on transmit-receive channels with I-flex topology when SDL or SDD used simultaneously with ABS topology
- Report generator does not support high quantity of indications with screenshots
- Sharck Fillet Weld Probe - Transverse C-scan displays too many channels
- Cursor indicating current position of probe does not display in C-Scan view during acquisition
- Magnifi crashes when applying a modification to the C-scan Derivate process
- Software update feature from the backstage can allow updates to a version not accessible to the user when a licence includes access to Magnifi 3 and 4
- Code view does not display C-scan saturation information
- C-scan subtraction cursor not properly working in some cases
- Encoded tubing array scans contain display and synchronisation between views issues (code view scrolling, slew, etc.)